



## MONTHLY THERMOMETER CHECK

Month	Jan	Feb	March	April	May	June	July	Aug	Sept	Oct	Nov	Dec
Reading in Ice Water												
Reading in Boiling Water												
Checked By (initial)												

The readings in **ice water** should be between **30°F (-1°C)**, and **34°F (1°C)**. If outside this range, the unit should be replaced or returned to the manufacturer to be recalibrated.

The readings in **boiling water** should be between **210°F (99°C)**, and **214°F (101°C)**. If outside this range, the unit should be replaced or returned to the manufacturer to be recalibrated.

Date of Yearly Calibration: \_\_\_\_\_ Date of Next Yearly Calibration: \_\_\_\_\_

Details/Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_